



Matcher Technologies Limited  
1B Canalside Business Park  
Tattenhall Chester  
Cheshire, CH3 9BD  
United Kingdom



140 Hale Street  
Haverhill, MA 01830  
[qc@embryotech.com](mailto:qc@embryotech.com)

ELI Accession Number: IIL-1927-0523

Date of completion: 05-19-2023

Lot number: 0202304A

Reference number: LBL020

Description of test article(s): Matcher Not For Cryo Use Labels

**Assay system requested by customer:** The test article (4) is placed on a culture plate. One-cell mouse embryos are placed in the culture plate and cultured for 96-hours.

**Control assay method and results:** 21 one-cell (B6C3F1 X B6D2F1) embryos were cultured in triplicate micro drops of culture medium overlaid with oil in control incubator (ELI-350):

21 / 21 (100 %)	1-cell to 2-cell within 24 hr
21 / 21 (100 %)	1-cell to expanded blastocyst within 96 hr

For a valid assay, Embryotech™ requires at least 80% of one-cell control embryos to develop to expanded blastocyst within 96-hours.

**Test assay method and results:** 21 one-cell (B6C3F1 X B6D2F1) embryos were cultured in triplicate micro drops of culture medium overlaid with oil with the test articles adhered to the outside of the culture plate in 350:

21 / 21 (100 %)	1-cell to 2-cell within 24 hr
21 / 21 (100 %)	1-cell to expanded blastocyst within 96 hr

Pass/Fail = Pass

**Summary of observations:** All test and control embryos were selected randomly from a common pool of freshly collected embryos. 100 percent of the control embryos developed to the expanded blastocyst stage within 96-hours. 100 percent of the test embryos cultured in the culture plate with the test articles adhered developed to the expanded blastocyst stage within 96-hours.

  
Signature  
Study Director

05.19.2023  
Date

  
Signature  
Quality Reviewer

05/19/23  
Date