



**Matcher Technologies Limited**  
**1B Canalside Business Park**  
**Tattenhall Chester**  
**Cheshire, CH3 9BD**  
**United Kingdom**

**EMBRYOTECH**  
LABORATORIES  
140 Hale Street  
Haverhill, MA 01830  
[gc@embryotech.com](mailto:gc@embryotech.com)

**ELI Accession Number:** IIL-4168-0524

**Date of completion:** 05-12-2024

**Lot number:** 0202404B

**Reference number:** LBL020

**Description of test article(s):** Matcher Not For Cryo Use Labels

**Assay system requested by customer:** The test article (4) is placed on a culture plate. One-cell mouse embryos are placed in the culture plate and cultured for 96-hours.

**Control assay method and results:** 21 one-cell (B6C3F1 X B6D2F1) embryos were cultured in triplicate micro drops of culture medium overlaid with oil in control incubator (ELI-248):

21 / 21 (100 %)	1-cell to 2-cell within 24 hr
21 / 21 (100 %)	1-cell to expanded blastocyst within 96 hr

*For a valid assay, Embryotech™ requires at least 80% of one-cell control embryos to develop to expanded blastocyst within 96-hours.*

**Test assay method and results:** 21 one-cell (B6C3F1 X B6D2F1) embryos were cultured in triplicate micro drops of culture medium overlaid with oil with the test articles adhered to the outside of the culture plate in ELI-346:

21 / 21 (100 %)	1-cell to 2-cell within 24 hr
21 / 21 (100 %)	1-cell to expanded blastocyst within 96 hr

**Pass/Fail = Pass**

**Summary of observations:** All test and control embryos were selected randomly from a common pool of freshly collected embryos. 100 percent of the control embryos developed to the expanded blastocyst stage within 96-hours. 100 percent of the test embryos cultured in the culture plate with the test articles adhered developed to the expanded blastocyst stage within 96-hours.

*CMFntw*  
Signature  
Study Director

*08-15-2024*  
Date

*Eric Lunn*  
Signature  
Quality Reviewer

*5/15/24*  
Date