



Matcher Technologies Limited
1B Canalside Business Park
Tattenhall Chester
Cheshire, CH3 9BD
United Kingdom

EMBRYOTECH
LABORATORIES
140 Hale Street
Haverhill, MA 01830
oc@embryotech.com

ELI Accession Number: IIL-6361-1125

Date of completion: 2025-11-17

Lot number: 0212510A

Reference number: LBL021

Description of test article(s): Matcher For Cryo Use Labels

Assay system requested by customer: The test article (2) is placed on a culture plate. One-cell mouse embryos are placed in the culture plate and cultured for 96-hours.

Control assay method and results: 21 one-cell (B6C3F1 X B6D2F1) embryos were cultured in triplicate micro drops of culture medium overlaid with oil in control incubator (ELI-336):

| | | | | |
|----|---|----|-------|--|
| 21 | / | 21 | 100 % | 1-cell to 2-cell within 24 hr |
| 21 | / | 21 | 100 % | 1-cell to expanded blastocyst within 96 hr |

For a valid assay, Embryotech™ requires at least 80% of one-cell control embryos to develop to expanded blastocyst within 96-hours.

Test assay method and results: 21 one-cell (B6C3F1 X B6D2F1) embryos were cultured in triplicate micro drops of culture medium overlaid with oil with the test articles adhered to the outside of the culture plate in incubator ELI-350:

| | | | | |
|----|---|----|-------|--|
| 21 | / | 21 | 100 % | 1-cell to 2-cell within 24 hr |
| 21 | / | 21 | 100 % | 1-cell to expanded blastocyst within 96 hr |

Pass/Fail = Pass

Summary of observations: All test and control embryos were selected randomly from a common pool of freshly collected embryos. 100 percent of the control embryos developed to the expanded blastocyst stage within 96-hours. 100 percent of the test embryos cultured in the culture plate with the test articles adhered developed to the expanded blastocyst stage within 96-hours.


Signature
Study Director


Date


Signature
Quality Reviewer


Date