



IMT International Ltd.
Canalside Business Park
Tattenhall Chester
Cheshire, CH3 9BD
United Kingdom

EMBRYOTECH
LABORATORIES
140 Hale Street
Haverhill, MA 01830
qc@embryotech.com

ELI Accession Number: IIL-7313-0817

Date of completion: 08-26-2017

Lot number: 0181706A

Reference number: LBL018

Description of test article(s): Matcher ID Security - For Cryo Use Label

Assay system requested by customer: The test article is placed on a culture plate. 1-cell mouse embryos are placed in the culture plate and cultured for 96-hours.

Control assay method and results: 15 1-cell (B6C3F1 X B6D2F1) embryos were cultured in triplicate micro drops of culture medium overlaid with oil in:

15 / 15 (100 %)	1-cell to 2-cell within 24 hr
15 / 15 (100 %)	1-cell to expanded blastocyst within 96 hr

For a valid assay, Embryotech™ requires at least 70% of 1-cell control embryos to develop to expanded blastocyst within 96-hours.

Test assay method and results: 21 1-cell (B6C3F1 X B6D2F1) embryos were cultured in triplicate micro drops of culture medium overlaid with the test article adhered to the outside of the culture plate:

21 / 21 (100 %)	1-cell to 2-cell within 24 hr
21 / 21 (100 %)	1-cell to expanded blastocyst within 96 hr

Summary of observations: All test and control embryos were selected randomly from a common pool of freshly collected embryos. 100 percent of the control embryos developed to the expanded blastocyst stage within 96-hours. 100 percent of the test embryos cultured in the culture plate with the test article adhered developed to the expanded blastocyst stage within 96-hours

A handwritten signature in blue ink.

signature
Study Director

08-28-2017
date

A handwritten signature in blue ink.

signature
Quality Reviewer

08-28-2017
date