

IMT International Ltd.  
Canalside Business Park  
Tattenhall Chester  
Cheshire, CH3 9BD  
United Kingdom

 EMBRYOTECH  
LABORATORIES  
140 Hale Street  
Haverhill, MA 01830  
[qc@embryotech.com](mailto:qc@embryotech.com)

ELI Accession Number: S3312-0318IIL

Date of completion: 03-29-2018

Lot number: N/A

Reference number: L7651


Description of test article(s): Avery Label Sheet

**Assay system requested by customer:** 1mL of sperm wash medium with sperm was added to a dish with the test article adhered to the outside and incubated for 24-hours. The forward progressive motility was read and recorded at 24-hours.


**Results:**

Test method: SOP/TSG/ELI/008	Specification	Initial	Result % 24hr	SMI Value	Pass/Fail
Test Article	SMI $\geq$ 0.75	98%	98%	1.00	Pass
Control	$\geq$ 70%	98%	98%	N/A	Pass

**Summary of observations:** All test and control sperm was prepared from the same donor and incubated in the same incubator at 32°C and 5% CO<sub>2</sub>. The control sperm had a 98% forward progressive motility at 24-hours. The test article sperm had a 98% forward progressive motility at 24-hours.

  
Signature  
Study Director

03-29-2018  
Date

  
Signature  
Quality Reviewer

03-29-2018  
Date